



**Direct Leisure Supplies Ltd, trading as DLS Medical**

Southgate Way  
Premier House  
Southgate Way, Orton Southgate  
PE2 6YG  
UK

11/09/2024

**Confirmation Letter Reference: CLNB1639 - GBPC 05933**

To whom it may concern,

Confirmation of receipt of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.

This letter confirms that, SGS Belgium NV, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number 1639 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

**Direct Leisure Supplies Ltd, trading as DLS Medical**

Southgate Way  
Premier House  
Southgate Way, Orton Southgate  
PE2 6YG  
UK  
SRN:

**Authorised Representative:**

MedEnvoy Global B.V  
Prinses Margrietplantsoen 33  
Suite 123  
2595AM The Hague  
The Netherlands  
SRN: NL-AR-000024028

The devices covered by the formal application and the written agreement mentioned above are identified in the Tables below . Table 1 identifies the devices which an MDR application has been

received, written agreement concluded and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but the NB has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that:

- The manufacturer signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry;
- The certificates expired after 26<sup>th</sup> May 2021 by course of time and were valid at the date of their expiry neither having been suspended nor withdrawn.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3 of MDR (as amended by EU 2023/607), are shown below:

- 26<sup>th</sup> May 2026 for Class III custom-made implantable devices
- 31<sup>st</sup> December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31<sup>st</sup> December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31<sup>st</sup> December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body SGS Belgium NV NB1639,

Virginie SILORET  
Global Medical Device Certification Manager  
Email: [Virginie.siloret@sgs.com](mailto:Virginie.siloret@sgs.com)  
Phone: +41 22 739 98 58

This document has been edited for confidentiality. If you require more detail of products covered in "Table 1" please contact us and we will forward a full copy of this letter.